

Addendum No.01

Minutes of the Pre Bid Meeting of Supply, Delivery, Installation and Commissioning of Laboratory Equipment for the Faculty of Medicine, Uva Wellassa University

Date: 12.09.2023
Time: 10.00 a.m.
Venue: Faculty of Medicine, Uva Wellassa University
Ref No: UWU/G/NCB/FOM/23/01

The Deputy Bursar (Stores and Supplies) welcomed the members of the TEC and bidders to the pre bid meeting. During the meeting following clarifications were notified to the bidders.

- Section II Bidding Data Sheet (BDS) ITB 35.4 clause deleted (Page 39).
- Technical Specifications

Package 01

- **1. Auto Clave (Bench Top) -**Specification No.17 deleted.
- **2. Digital Top Loading Balance –**
 - Specification No.01 Capacity shall not be less than 1000g.
 - Specification No.09 Level indicator for easy viewing.
 - Specification No.11 deleted.
- **3. Analytical Balance-**
 - Specification No.01 Capacity 60-120 g (should including the range)
 - Specification No.04 Linearity – at least ± 0.1 mg
- **5. Centrifuge machine (15ml and 50ml Tubes) –**
 - Specification No.02 The unit should have a maximum capacity of not less than 400ml and should be supplied with glass tubes to centrifuge at least 16 x 15ml standard tubes in each load.
 - Specification No.10 deleted.
- **6. Centrifuge machine (Eppendorf micro centrifuge Tubes) –**
 - Specification No.10 deleted.
- **8. Distillation water plant-**
 - Specification No.7 deleted.
- **9. Magnetic stirrers with hot plate –**
 - Specification No.15 Protection rating should be IP 42 or equivalent.

- **17. Water Bath (Boiling)-**
 - Changed as “Water Bath”.
 - Specification No.03 It should be available with temperature settings and functioning thermometer with thermostatic temperature with 5°C above ambient to 100°C.

Package 02

- **22. Refrigerator (Standard) –**
 - Specification No.15 deleted.

Package 03

- **24. Centrifuge –Micro hematocrit centrifuge –**
 - Specification No.01 Speed not less than 10,000 rpm step of 100
 - Specification No.03 Capillary tube.
 - Specification No.09 deleted.
- **25. Centrifuge –Ordinary electric**
 - Specification No.02 Specimen capacity: 12 tube/run or higher.
 - Specification No.03 Sample size: 24 x 10 ml wintrobe tube (or equivalent) should be supplied.
 - Specification No.09 deleted.
- **34. Weighing Scales**
 - Specification No.11 The scale should be as per BIS/ISO standard.
- **36. Osmo meter**
 - Specification No.02 Sample volume less than 1ml.

Package 09

- **55. Light Microscope (binocular)**
 - Specification No.01 The equipment shall be suitable for use on AC mains 230V ± 10%, 50 Hz.
 - Specification -General Conditions - No.10 deleted.



Deputy Bursar
(Stores & Supplies)
Uva Wellassa University

Deputy Bursar
Stores & Supplies
Uva Wellassa University
Passara Road, Baa