

MINUTES OF PRE-BID MEETING

Supply, Delivery and Installation of Laboratory and Teaching Equipment for The Year – 2026,
Faculty of Animal Science and Export Agriculture, Uva Wellassa University

Tender no : UWU/G/NCB/C/26/06

Date & Time : 13/05/2026 & 2.00 PM

Venue : CODL Board Room

Present Bidders

- a. Mr. V.T.M.Amashanth – Hemson International (Pvt)Ltd – 0772992051 / thilina@hemson.lk
- b. Mr. S.R.Mannapperuma – Microtech Biological (Pvt) Ltd- 0702635042 / sales3@microtechbio.com
- c. Mr. Chathuranga – CMC Engineering Export GmbH – 0762872592/ chathuranga@cmcenglk.com
- d. Mr.Sameera – Avon Pharmo Chem (Pvt) Ltd – (Joined the meeting through Zoom)
- e. Mr Malith - Avon Pharmo Chem (Pvt) Ltd – (Joined the meeting through Zoom)

At the Pre-bid meeting bidders were requested the amendment to the specifications of following items. Bid Evaluation Committee decided to amend the specifications as follows as per the clarifications of bidders.

Item No	Description	Remark
02	Chest Freezer (Curved Glass Top) - Domestic Use	Attached the amended specification with bidding document
03	Chest Freezer - Domestic Use	
10	Two door side by side refrigerator with freezer - Domestic Use	
11	Trinocular stereo microscope with camera (Dissecting Microscope) - Warranty: Minimum 05 years	
12	Work plate size : Capacity -50ml – 2000ml	
14	Vortex mixer No 1: Time Setting: Timing display	
18	Binocular Microscope - Minimum 05 years	
21	Rotary Evaporator - Tank volume: 3.5 L–4.00 L Refrigerant: 134A or equivalent Protection class [Ingress Protection]: DIN EN 60529-IP 20 or equivalent Instrument should be European CE Certified or equivalent	

24	Dual Purpose Laminar Flow Front Panel – Adjustable Work Area Width more than (mm) - 1400 mm	
26	Vortex Mixture No 2 Variable speed control: Yes Interchangeable mixing heads: Yes	
27	Autoclave with dryer Sterilizing Temperature: Maximum temperature should be 140°C Power Supply: Three phase	
36	Drying Oven Resistance Wire Material: (Nickel-Chromium alloy (NiCr 80/20) or equivalent	
41	Acclimatization Chamber Door: Lockable Double door	

Chairman
Department of Procurement Committee(Major)
Uva Wellassa University